



# RBF Morph


Enabling Medical Digital Twin through Advanced Mesh Morphing and High-Fidelity Patient-Specific Simulations

Marco Evangelos Biancolini  
Company founder @RBF  
Associate Professor @UTV




# We make CAE models parametric

- RBF Morph makes the CAE model parametric
- Shape parameters are driven by an orchestrator
- Shape parameters can be used to generate snapshots for real time Digital Twins (**ROM/AI**)




## RBF's morphing approach



# Radial Basis Functions mesh Morphing

- Geometric control by **Radial Basis Functions mesh Morphing**
  - Surface shape changes
  - Volume mesh adaption
- A **new shape** of the CAE model **ready to run**
  - for structures in the FEA solver
  - for flows in the CFD solver



# Radial Basis Functions mesh Morphing

- We offer **Radial Basis Functions** (RBF) to drive mesh morphing (smoothing) from a list of source points and their displacements











- RBF are recognized to be one of the **best mathematical tool** for mesh morphing

$$\begin{cases} s_x(\mathbf{x}) = \sum_{i=1}^N \gamma_i^x \varphi(\|\mathbf{x} - \mathbf{x}_{s_i}\|) + \beta_1^x + \beta_2^x x + \beta_3^x y + \beta_4^x z \\ s_y(\mathbf{x}) = \sum_{i=1}^N \gamma_i^y \varphi(\|\mathbf{x} - \mathbf{x}_{s_i}\|) + \beta_1^y + \beta_2^y x + \beta_3^y y + \beta_4^y z \\ s_z(\mathbf{x}) = \sum_{i=1}^N \gamma_i^z \varphi(\|\mathbf{x} - \mathbf{x}_{s_i}\|) + \beta_1^z + \beta_2^z x + \beta_3^z y + \beta_4^z z \end{cases}$$

# Main uses of RBF Morph


Usage	Mechanical	Fluent	optiSLang	Twin Builder
Automated and quick variable design space exploration.	✓	✓		
Optimization (Single physics or multi-physics). Shape optimization for stress reduction, mass reduction, fluid-structure interaction	✓	✓	✓	
Digital twin development (static ROMs)	✓	✓	✓	✓
Lifing applications Simulate defects such as corrosion pits, spalling of material, erosion, chips, etc.	✓	✓		
Examine the effects of non-conformance and manufacturing variability	✓	✓		
Robust Design	✓	✓	✓	

# Applications



# EU-funded research projects

**RBF4AERO**



R I B E S




EUROfusion



MEDITATE



# MeDiTATe Endovascular Abdominal Aneurysm Repair



ST. OLAVS HOSPITAL  
UNIVERSITETSSYKEHUSET I TRONDHEIM



**Ansys**  
LS - DYNA




**Ansys**  
TWIN BUILDER


**Ansys**  
TWIN BUILDER



# MeDiTATe Endovascular Abdominal Aneurysm Repair




# Medical Digital Twin Copernicus



[https://www.ff4eurohpc.eu/en/experiments/2021070910512579/cloudbased\\_hpc\\_platform\\_to\\_support\\_systemicpulmonary\\_shunting\\_procedures](https://www.ff4eurohpc.eu/en/experiments/2021070910512579/cloudbased_hpc_platform_to_support_systemicpulmonary_shunting_procedures)

# Medical Digital Twin Copernicus



# Medical Digital Twin Copernicus



Working Directory: .../EKA01/Desktop/rbp-output-FINAL/velocity-magnitude | 150 Snapshots - 12 Parameters | Browse | Set Geometry ROM | Auto

Check | Build | Validate | Evaluate

**Available ROM:**

- velocity

Delete | Rename

**ROM Information**

Name : velocity  
Parameters : 12  
Learning Snapshots : 120  
Modes : 18  
Version : 2022R1

Export to Twin builder

**Evaluate Roms**

Input parameters :

Parameter	Value
dL1_vol	-1.50000e-01
dL2_vol	-3.40000e-01
dr1_vol	4.90000e+00

Points Size: 0.00 | Views: | Parts: all | CutPlane 0

**velocity**

Velocity Magnitude (m/s)

The 3D visualization shows a complex, branching structure, likely representing a blood vessel or airway. The color scale indicates velocity magnitude in m/s, with the lowest values (blue) at the base and higher values (red) at the top branches.


# Medical Digital Twin DiTAiD



## From lung scan to medical use




1) Scan of lungs



2) Extraction of lung shape parameters



3) Digital twin




4) Visualization and interpretation for medical use



# Medical Digital Twin DiTAiD



- Base geometry is obtained from literature
  - ✓ Constructed from several high-resolution CT scans of 47 year old healthy volunteer
  - ✓ The base geometry has been studied in multiple experimental and numerical studies
  - ✓ Includes up to the 4<sup>th</sup> generation (note, human lungs go up to 23 generations)
- Identify relevant input parameters for the digital twin
  - ✓ Shape
  - ✓ Flow
  - ✓ Particle





Z. Zhang, C. Kleinstreuer and S. Hyun, "Size-change and deposition of conventional and composite cigarette smoke particles during inhalation in a subject-specific airway model," *Journal of Aerosol Science*, vol. 46, pp. 34-52, 2012.

S. Kenjereš and J. L. Tjin, "Numerical simulations of targeted delivery of magnetic drug aerosols in the human upper and central respiratory system: a validation study," *Royal Society Open Science*, vol. 4, no. 12, p. 170873, 2017.

# Medical Digital Twin DiTAiD



- Potentially a huge amount of shape parameters!
- Amount of input parameters is limited by assuming:
  - ✓ Circularity is kept constant
  - ✓ Only considered angle is the branching angle
  - ✓ Diameter follows a fixed ratio of  $h=0.79$
- Mouth-throat part: 3 parameters
- Lower airways: 23 parameters
  - ✓ Generation 0 (trachea): 1L, 1D, 1A
  - ✓ Generation 1: 2L, 2A
  - ✓ Generation 2: 4L, 4A
  - ✓ Generation 3: 8L



T. Van de Moortele et al.; "Morphological and functional properties of the conducting human airways investigated by in vivo computed tomography and in vitro MRI"

Generation	Diameter [mm]	Length [mm]		Branching angle [deg]
		Left	Right	
<b>0 (Trachea)</b>	15 - 20	100 - 120		80 - 95
<b>1</b>		51 - 57	24 - 28	75 - 90
<b>2</b>		12 - 16	15 - 28	65 - 95
<b>3</b>		7 - 10	7 - 10	55 - 70

# Medical Digital Twin DiTAiD



<https://www.flickr.com/photos/aceofknives/25604600281/>

Physical parameters: 3 parameter

- ✓ Flow rate varies between 15 L/min and 120 L/min
- ✓ Particle size varies between 0.1  $\mu\text{m}$  and 10  $\mu\text{m}$
- ✓ Particle injection rate varies between 0 m/s and 10 m/s

26 shape parameters and 3 physical parameter

29 input parameters in total







# Medical Digital Twin DiTAiD




- Design Of Experiments (DOE) table is generated:
  - ✓ For the 29 input parameters
  - ✓ Using the Latin Hypercube Sampling for optimal spacing
  - ✓ Creating 1000 design points
- Fluent settings validated in literature
  - ✓ Steady state
  - ✓ RANS, transitional SST (4eq)
  - ✓ Particles are one-way coupled




# Medical Digital Twin DiTAiD



# Medical Digital Twin DiTAiD



# Medical Digital Twin DiTAiD



# Ansys RBF Morph products

- An RBF mesh morphing solution fully embedded in Ansys
  - RBF Morph Fluids – an Add On for Fluent
  - RBF Morph Structures – an ACT App for Mechanical
- Full integration with **optiSLang** and **Twin Builder**
- Support for **LS-DYNA** and **APDL**



**Add-On Packages**

Ansys RBF Morph ?


↓ Fluids

↓ Structures

[https://www.rbf-morph.com/wp-content/uploads/2023/05/RBFMorph\\_Brochure.pdf](https://www.rbf-morph.com/wp-content/uploads/2023/05/RBFMorph_Brochure.pdf)

# New RBF Morph Stand Alone


- To be released in 2024
- Read in STL, STEP
- Unity - OpenCascade
- Solver independent process that supports many mesh formats
- Scriptable via python



# Next step?

A complete solution to deliver interactive digital twins with AR/VR custom UI


- FMU are translated to ARM
- Meta Quest 3
- Input parameters are controlled by hands



# Conclusions



- Medical Digital Twins are feasible today!
- The **In Silico** path, i.e. MDT driven by high fidelity simulations, is ready and requires
  - Patient specific data (from images)
  - State of the art multi-physics simulation
  - Reduced order models and advanced mesh morphing
- A clear **business model** is required
  - Public funds are today the major resource
  - Certification is complex
- We are moving in the right direction and there is **mainstream focus** on Medical Digital Twins





# Thank you!

marco.biancolini@rbf-morph.com



[linkedin.com/company/rbf-morph](https://www.linkedin.com/company/rbf-morph)



[youtube.com/user/RbfMorph](https://www.youtube.com/user/RbfMorph)



[rbf-morph.com](https://www.rbf-morph.com)